



Cipla

Bedometasone and Salbutamol Inhaler

Aerocort-A HFA inhaler

with **dose indicator**


patient information leaflet

please read this leaflet completely before use



ABOUT YOUR AEROCORT-A HFA INHALER


PARTS OF THE INHALER



Your **AEROCORT-A HFA** inhaler now comes with a dose indicator. It shows the number of puffs in the inhaler. As you use the inhaler, the dose indicator will countdown and indicate the number of puffs remaining.

HOW TO KNOW THAT YOUR AEROCORT-A HFA INHALER IS GETTING OVER


When there are 40 puffs remaining, the colour of the numbers will change from green to red.



This indicates that fewer doses are remaining in the inhaler. You should now consider getting a new inhaler or ask your doctor if you need another one.

When the dose indicator displays '0', this means that there is no more medicine left in the inhaler & you need to

discard the inhaler. Your inhaler may not feel empty & it may continue to operate, but you will not get the right amount of medicine, if you keep using it beyond '0'.



BEFORE USING YOUR AEROCORT-A HFA INHALER

- A.** Remove the cap from the mouthpiece & make sure that the mouthpiece is clean.
- B.** Hold the inhaler away from your face. Shake it well & release two puffs into the air.



C. Your **AEROCORT-A HFA** inhaler is now ready for use. **IF you have not used your inhaler for a week or more, shake well and release one puff into the air.**

USING YOUR AEROCORT-A HFA INHALER

1. Sit or stand upright. Remove the mouthpiece cap & shake the inhaler well. Hold it upright as shown, with your thumb at the base below the mouthpiece. Place either one or two fingers on top of the canister.



2. Breathe out fully, through your mouth.




3. Place the mouthpiece of the inhaler in your mouth between your teeth & close your lips around it (do not bite it). **Start breathing in slowly through your mouth. Press down the canister firmly & fully to release one spray while continuing to breathe in slowly & deeply.**



4. Remove the inhaler from your mouth & hold your breath for 10 seconds, or for as long as is comfortable. Breathe out normally.




5. If another puff is required, wait for at least 1 minute. Shake inhaler well & repeat steps 2 to 4. After use, replace the mouthpiece cap firmly & snap it into position.




6. After taking each dose, rinse your mouth with water & spit it out.

IMPORTANT: Do not rush steps 2, 3 & 4. It is important that you start to breathe in slowly before releasing a puff. To ensure correct use of the

inhaler, use it in front of a mirror for the first few times. If you see 'mist' coming out from the top of the inhaler or the sides of your mouth, start again from step 1. This escaping mist indicates incorrect technique.




In case of difficulty in using the inhaler correctly, you may use it along with a Zerostat-VT spacer.



CLEANING YOUR AEROCORT-A HFA INHALER

It is important to keep your inhaler clean. Clean your inhaler atleast once a week.

- 1.** Take the mouthpiece cap off. **DO NOT take the metal canister out of the actuator.**
- 2.** Wipe the inside & the outside of the mouthpiece with a clean, dry cloth.



- 3.** Replace the mouthpiece cap.
- 4. DO NOT wash or soak any part of the inhaler in water.**

STORING YOUR AEROCORT-A HFA INHALER

Store below 30°C. Avoid storage in direct sunlight or heat. Protect from frost. Keep the inhaler in an upright position, with the mouthpiece down.

DO NOT

- × Spray the inhaler in your eyes.
 - × Exceed the recommended dose.
 - × Change/tamper with the numbers on the dose indicator.
 - × Puncture or burn the inhaler even when empty as it is pressurized.
- Keep the inhaler out of the reach of children.



21070341



PACKAGING DEVELOPMENT

Product Name: Aerocort-A HFA Inhaler		Material No.: 21070341	Version: 01	Item: Leaflet	Co-ordinator: Shilpa	Artist: Manish	Date: 23/05/2018
Colours: BLUE WOOL TEST VALUE 5-9 (LIGHT FASTENING DATA)		Front Side ■ Pantone 5473 C ■ Pantone 425 C ■ Pantone 185 C ■ Pantone 354 C	Back Side ■ Pantone 425 C		INK: Oil based Ink from DIC OR MICRO		
Design: Unfolded		Reference: 21048078		Software: Illustrator CC			
Fonts: —		Links: Aerocort Inhaler.tif					
Actual Size: 280 x 170 mm	Size after folding: NA	2D Code: 21070341(F) & 21070341(B)		Grain Direction : Parallel to length		Screen : # __	
Material: 70 GSM Maplitho Paper (From Andhra OR Ballarpur Paper Mill ONLY)				Varnish:		Artwork Print Size: <input type="checkbox"/> actual <input type="checkbox"/> scaled	
Path: D:\Manish\April 2018 to March 2019\Shilpa\Lebanon\Aerocort A Inhaler\21070341 Aerocort-A HFA Inhaler Dose Indicator Leaflet Lebanon.ai							
<ul style="list-style-type: none"> • Instructions / Remark: NA • Any deviation must be brought to the notice of packaging development co-ordinator immediately. • For any clarification, please contact packaging development co-ordinator immediately. • NO CHANGES IN ARTWORK SHOULD BE DONE BY THE PRINTER • The printer should verify the e-proof against the approved artwork before submitting for approval and the e-proof should have printer details. 		Checked by	Artist	Cordinator	file loaded in Server	Section Head	
Pharma Code		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
2D Code		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
QR Code		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Bar Code		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Artwork		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Spell check		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	